

Ammonia-Nitrogen

0-2 ppm

C-3304

-
1. Fill the test tube with the water sample to the 5 ml mark.
 2. Add 10 drops of Salicylate Ammonia #1. Cap and mix.
 3. Add 7 drops of Salicylate Ammonia #2. Cap and mix. Wait 1 minute.
 4. Add 7 drops of Salicylate Ammonia #3. Cap and mix. Wait 20 minutes
 5. Insert the Ammonia-Nitrogen Octa Slide Viewer, insert the test tube into the viewer and match the color. Record as ppm of Ammonia-Nitrogen (N-NH₃)

To convert ppm of Ammonia-Nitrogen into ppm of Ammonia (NH₃), multiply the above reading by 1.2

Replacement Reagents and Parts

R-3978LWT/ 2oz	-	Salicylate Ammonia #1, with nasal plug
R-3979WT/ 1oz	-	Salicylate Ammonia #2, with nasal plug
R-3982WT/ 1oz	-	Salicylate Ammonia #3, with nasal plug
P-3441	-	Comparator slide viewer, Ammonia-Nitrogen 0-2 ppm
P-1100	-	Slide viewer
P-0106 (x2)	-	Graduated test tube w/cap, plastic, 2.5, 5 and 10 ml