

CHLORIDE/SODIUM CHLORIDEK-0350/10

1. Fill a casserole or Erlenmeyer flask with 50 ml of water sample.
2. Add 15 drops of potassium chromate indicator solution. The water will turn yellow.
3. Add Chloride titrating solution until a faint brick red color appears throughout the entire sample.

Cl ppm = ml of titrating solution x 10

NaCl ppm = Cl ppm x 1.65

Replacement Reagents and Parts

R-2025/qt/10	-	Silver nitrate titrating solution (1 ml = 10 ppm Cl/50 ml)
R-1080/2oz	-	Potassium chromate indicator solution w/nasal plug
P-1088	-	140 ml Porcelain Casserole
P-1140	-	Automatic buret 10 ml bead type w/500ml reservoir